

Drug Recall

DATE OF RECALL: June 10, 2024

DRUG NAME: Accu-Chek Guide Kit

RECALLING FIRM: Roche

REASON FOR RECALL: Roche announced a user-level recall of some Accu-Chek Guide Kits because they were configured incorrectly and show results in mmol/L rather than in the labeled measuring unit, mg/dL.

FDA LINK: N/A

OTHER DETAILS: N/A

RECALLED PRODUCT:

Product Description	NDC Number	Lot Number	Expiration Date
Accu-Chek Guide Kit	65702072910	92339920445	08/01/23 06/10/24
		92339955415	
		92340116052	
		92340117408	
		92340120006	
		92340120057	
		92339920116	
		92339094787	
		92339744998	