

Drug Recall

DATE OF RECALL: February 13, 2023

DRUG NAME: BD Insulin Syringes with the BD Micro-Fine™ IV Needle

RECALLING FIRM: Embecta Corp.

REASON FOR RECALL: This recall was issued due to using unvalidated materials that can result in unsealed packaging for the individual syringes.

FDA LINK: N/A

OTHER DETAILS: N/A

RECALLED PRODUCT:

| Product Description | NDC Number | Lot Number | Expiration Date |
|--|-----------------|------------|-----------------|
| BD Insulin Syringes with the BD Micro-Fine™ IV Needle 1mL, 0.35mm (28G) x 12.7mm (1/2") U-100, 100 count | 08290-3294-20** | 2038204 | 4/30/2027 |
| BD Insulin Syringes with the BD Micro-Fine™ IV Needle 1mL, 0.35mm (28G) x 12.7mm (1/2") U-100, 100 count | 08290-3294-24 | 2024074 | 4/30/2027 |
| BD Insulin Syringes with the BD Micro-Fine™ IV Needle 0.5mL, 0.35mm (28G) x 12.7mm (1/2") U-100, 100 count | 08290-3294-61** | 2031528 | 4/30/2027 |