

## Drug Recall

**DATE OF RECALL:** June 26, 2023

**DRUG NAME:** CONTOUR® NEXT Gen Blood Glucose Meters

**RECALLING FIRM:** Ascensia Diabetes Care.

**REASON FOR RECALL:** This recall was issued due to meters distributed in the United States packaged into the meter kit having incorrect factory-set units of measurement where the meters display glucose results in mmol/L rather than mg/dL.

**FDA LINK:** N/A

**OTHER DETAILS:** N/A

**RECALLED PRODUCT:**

Product Description	NDC Number	Lot Number	Expiration Date
Asthmanefrin Racpinephrine Inhalation Solution Bronchodilator (racephedrine, USP 11.25), 30x0.5 mL Sterile Vials	00487-2784-01	126451	04/30/2023
S2 Racpinephrine Inhalation Solution, USP 2.25% Bronchodilator 30x0.5 mL	00487-5901-99	126461 126431	04/30/2023