

## Drug Recall

**DATE OF RECALL:** July 31, 2024

**DRUG NAME:** FreeStyle Libre 3 Sensors

**RECALLING FIRM:** Abbott.

**REASON FOR RECALL:** Abbott announced a consumer level recall of three lots of FreeStyle Libre 3 sensors because the sensors may provide incorrect high glucose readings.

**FDA LINK:** N/A

**OTHER DETAILS:** N/A

**RECALLED PRODUCT:**

Product Description	NDC Number	Lot Number	Expiration Date
FreeStyle Libre 3 Sensors	57599-0818-00	T60001948 T60001966 T60001969	N/A