

Drug Recall

DATE OF RECALL: January 20, 2026

DRUG NAME: MediHoney wound and burn products

RECALLING FIRM: Integra Lifesciences

REASON FOR RECALL: Integra Lifesciences announced a voluntary, consumer-level recall of all unexpired lots of some MediHoney wound and burn products because packaging failures were identified which could lead to a breach in the sterile barrier.

FDA LINK: N/A

OTHER DETAILS: N/A

RECALLED PRODUCT:

Product Description	NDC Number	Lot Number	Expiration Date
MediHoney wound and burn products	09958003131	N/A	N/A
	09958003161		
	09958003471		
	09958003141		
	09958003461		