

Drug Recall

DATE OF RECALL: May 5, 2023

DRUG NAME: Medical Device Removal

RECALLING FIRM: Roche Diagnostic Corporation regarding Pilot™.

REASON FOR RECALL: This medical device removal was issued due to potentially harmful bacteria have been found in the liquid buffer solution of the kits.

FDA LINK: N/A

OTHER DETAILS: N/A

RECALLED PRODUCT:

Product Description	NDC Number	Lot Number	Expiration Date
Medical Device Removal Notification: Pilot COVID-19 At-Home Test	87473-0000-20	53K38N1T1	N/A
		53K38N2T1	
		53K38N3T1	
		53K38N4T1	
		53K38N5T1	
		53K38P1T1	
		53K38P2T1	
		53K38P3T1	
		53K41T5T1	
		53K41X1T1	
		53K41X2T1	
		53K41X3T1	
		53K4211T1	
		53K4212T1	
		53K4213T1	
		53K4221T1	
		53K4222T1	
		53K4223T1	
		53K4224T1	
		53K4225T1	
53K4231T1			
53K4232T1			
53K4233T1			

		53K4261T1 53K4262T1 53K4271T1 53K4272T1 53K4273T1 53K4274T1 53K4291T1 53K4292T1 53K42A1T1 53K42A2T1 53K42A3T1 53K42E1T1 53K42G1T1 53K42G2T1 53K42H1T1 53K42H2T1 53K42L1T1 53K42L2T1 53K4361AC 53K4362AC 53K4392AC	
--	--	---	--

