

Drug Recall

DATE OF RECALL: May 26, 2026

DRUG NAME: Omnipod® 5 and Omnipod DASH® Pods

RECALLING FIRM: Insulet

REASON FOR RECALL: On May 26, 2026, Insulet announced a voluntary medical device correction of some lots of Omnipod 5 and Omnipod DASH Pods because there may be a small tear in the tubing causing insulin to leak inside the Pod instead of going into the body. This can lead to untreated high blood sugar levels for a long time, which may lead to serious health problems. This correction is separate from the voluntary correction for Omnipod 5 Pods issued on March 12, 2026.

FDA LINK: [Insulet Initiates Voluntary Medical Device Correction for Certain Omnipod® Pods in the U.S. and Affected International Markets | FDA](#)

OTHER DETAILS: N/A

RECALLED PRODUCT:

Product Description	NDC Number	Lot Number	Seq No
Omnipod 5 Pods Omnipod DASH Pods	All NDCs	Contact omnipod.com/check- pods to see if patient's lot is affected	Ex: 1234567